

Patient ID	Scanner Manufacturer	Scanner Model	Field Strength	Sequence	Repetition Time (ms)	Echo Time (ms)	Slice Thickness (mm)	Number of Slices
ADEM01	Siemens	Symphony	1,5	MPR	1910.0	4.2	1.5	104
ADEM02	Philips	Achieva	1.5	T13DKM	25.0	4.6	1.6	180
ADEM03	Siemens	Avanto	1.5	T1MPRTRA	1160.0	4.1	0.9	192
ADEM04	Siemens	Verio	3	T2TSETRA448	4980.0	92	5.0	25
ADEM05	Siemens	Symphony	1.5	MP3DCOR	1830.0	3.6	1.2	165
ADEM06	Philips	Achieva	1.5	T1FFE3D	25.0	4.6	1.6	180
ADEM07	Philips	Achieva	1.5	T1FFE3DKM	25.0	4.6	1.6	190
ADEM09	Philips	Achieva	1.5	ST1W3DISOSAG	25.0	4.6	1.0	360
ADEM10	Siemens	Avanto	1.5	T1SETRA320	500.0	9.8	5.0	24
ADEM11	Siemens	Symphony	1.5	T1MPRSAG	1760.0	3.1	1.0	144
ADEM12	Siemens	Avanto	1.5	T1MPRWESAGKM	2040.0	3.0	1.0	160
ADEM13	Siemens	Avanto	1.5	T1SETRA	359.0	17.0	5.0	26
ADEM14	Philips	Achieva	1.5	FFET1COR	214.7	2.3	6.0	28
ADEM16	Siemens	Symphony	1.5	T1MPRNSTRA	1710.0	4.4	1.2	128
ADEM17	Siemens	Symphony	1.5	T1MPRAXKAM	1780.0	4.4	1.2	140
ADEM18	Siemens	Avanto	1.5	T1SETRA	562.0	13.0	5.0	25
ADEM19	Philips	Intera	1.5	T2SETRA	6899.0	110.0	4.0	31
ADEM20	Siemens	Verio	3	T1MPRTRA	1600.0	3.4	1.2	104
ADEM21	Siemens	Sonata	1.5	T1MPRNSTRA	1940.0	4.4	1.2	128
ADEM22	Siemens	Avanto	1.5	T1MPRTRAISO	1900.0	3.4	1.0	160
ADEM23	GE Medical	Signa	1.5	T2FRFSEAX	5050.0	101.6	5.0	24
ADEM24	Siemens	Trio	3	SWI	27.0	20.0	2.0	72
ADEM26	Philips	Ingenia	3	T1WFFESENSE	367.8	4.6	4.0	34
ADEM27	GE Medical	Discovery	3	T2FRFSEAX	11000.0	111.8	3.0	44
ADEM28	Siemens	Avanto	1.5	T2SETRA	5471.6	113.0	5.0	27
ADEM30	Philips	Achieva	3	T1FFE3DKM	20.0	4.6	1.0	308
ADEM31	Siemens	Sykra	3	T1MPRSAG	1880.0	2.7	1.0	160
ADEM32	Siemens	Symphony	1.5	T1MPRTRA	1830.0	3.6	1.2	128
ADEM33	Philips	Achieva	3	T1W3DKMTFE	8.2	3.8	1.0	163
ADEM34	Siemens	Symphony	1.5	T1MPRTRA	1830.0	3.6	1.2	128
ADEM35	GE Medical	Signa	1.5	T2AX	2516.7	100.9	5.0	30
ADEM36	Philips	Panorama	1	T1WFFE	177.3	2.1	5.0	25
ADEM37	Philips	Gyrosan	1.5	T2AX	6167.3	100.0	5.0	30
ADEM38	Philips	Achieva	3	T13DTFESAGKM	8.6	4.0	1.0	191
ADEM39	Siemens	Sonata	1.5	T1MPRSAG	1900.0	4.4	1.3	128
ADEM40	Siemens	Avanto	1.5	T1SETRA	562.0	13.0	5.0	25
ADEM41	Philips	Ingenia	3	T1AX	8.0	3.0	1.5	112
ADEM42	Hitachi	Echelon	1.5	T2AX	4159.0	96.0	5.0	22
ADEM43	Siemens	Symphony	1.5	T1TIRMTRA	8700.0	85.0	5.0	27
ADEM44	Siemens	Symphony	1.5	T1MPRNSTRA	1710.0	4.4	1.2	128
ADEM45	Siemens	Symphony	1.5	T1MPRAX	1860.0	3.6	1.0	176
ADEM46	Philips	Gyrosan	1.5	T1TRA	350.0	78.0	3.0	44
ADEM47	Siemens	Espreo	1.5	T1SETRA	624.0	11.0	5.0	24
ADEM48	GE Medical	Signa	1.5	T1AX3D	12.5	5.3	1.4	208

ADEM49	Siemens	Symphony	1.5	T2SETRA	6790.0	104.0	2.0	77
ADEM50	Philips	Intera	1.5	ST1W3DTFESAG	7.2	3.3	1.0	160
ADEM51	Siemens	Sonata	1.5	T1MPRSAG	1420.0	4.8	1.6	96
ADEM52	Siemens	Trio	3	T2TRADF	9000.0	134.0	3.0	42
ADEM53	Siemens	Symphony	1.5	T1SETRA	485.0	14.0	6.0	21
ADEM54	Siemens	Skyra	3	T1MPRAGESAG	2300.0	2.3	0.9	176
ADEM55	GE Medical	Discovery	3	T1MEMPSEAX	640.0	18.0	4.0	36
ADEM56	GE Medical	Signa	1.5	T1AX3DKM	10.3	4.3	1.2	116
ADEM57	Siemens	Avanto	1.5	T1MPRTRAIISO	1850.0	3.5	0.8	176
ADEM58	Siemens	Avanto	1.5	T1SETRAKM	500.0	8.1	5.0	24
ADEM59	Philips	Panorama	1	T2TSECOR	5687.2	110.0	5.0	24
ADEM60	Siemens	Avanto	1.5	T1MPRTRAIISO	1900.0	3.4	1.0	176
ADEM61	Siemens	Avanto	1.5	T1MPRTRAIISO	1900.0	3.4	1.0	176
ADEM61	Philips	Gyrosan	1.5	T2TSETRA	4773.7	110.0	5.0	18